

CERTIFICATE

SABS WATER CHECK PROFICIENCY TESTING

SANAS accredited, reference no PTS 0003, according to the requirements of ISO/IEC 17043: 2010 "General requirements for proficiency testing providers"

Set Point Laboratory - ISANDO (R)

STUDIES ENROLLED FOR JAN 2022 – DEC 2022

Group 1 – Metals

Group 2 – Nutrients

Group 3 – Minerals

$Z < 2$ = Satisfactory

$2 < Z < 3$ = Questionable

$Z > 3$ = Unsatisfactory

Note 1: Participating laboratory should investigate Z-score outliers according to their Laboratory quality system

R= amended

MONTH	GROUP	ABSOLUTE AVERAGE Z-SCORE	NUMBER OF TESTS PERFORMED	SATISFACTORY/ UNSATISFACTORY
2022 JANUARY	1	2.39	102	Questionable
2022 APRIL		-	-	-
2022 JULY		1.03	63	Satisfactory
2022 OCTOBER		-	-	-
2022 FEBRUARY	2	-	-	-
2022 MAY		1.81	14	Satisfactory
2022 AUGUST		-	-	-
2022 NOVEMBER		0.99	15	Satisfactory
2022 MARCH	3	0.86	31	Satisfactory
2022 JUNE		-	-	-
2022 SEPTEMBER		1.49	29	Satisfactory
2022 DECEMBER		-	-	-

C Modika 
Scheme Manager: Technical Signatory

MF Maleka 
Coordinator: Technical Signatory

SABS COMMERCIAL SOC Ltd. Reg. No. 2000/013581/30

Website: www.sabs.co.za E-mail: info@sabs.co.za Call Centre: 086 1277 227



PTS0003

Gauteng Head Office

1 Dr Lategan Road, Groenkloof
Private Bag X191
Pretoria, 0001
Tel +27 (0) 12 428 7911
Fax +27 (0) 12 344 1568

1233