

CERTIFICATE

SABS Water Check Proficiency Testing

SANAS accredited, reference no PTS 0003, according to the requirements of ISO/IEC 17043: 2010 "General requirements for proficiency testing providers"

Set Point Laboratory - Cape Town

STUDIES ENROLLED FOR JAN 2019 – DEC 2019

Group 1 – Metals
Group 2 – Nutrients
Group 3 – Minerals

Z < 2 = Satisfactory
2 < Z < 3 = Questionable
Z > 3 = Unsatisfactory

MONTH	GROUP	ABSOLUTE AVERAGE Z-SCORE	NUMBER OF TESTS PERFORMED
2019 JANUARY	1	-	-
2019 APRIL		-	-
2019 JULY		0.60	54
2019 OCTOBER		0.74	54
2019 FEBRUARY	2	-	-
2019 MAY		-	-
2019 AUGUST		1.34	14
2019 NOVEMBER		2.56	16
2019 MARCH	3	-	-
2019 JUNE		1.31	36
2019 SEPTEMBER		0.82	34
2019 DECEMBER		0.73	37

C Modika 
Scheme Manager: Technical Signatory

MF Maleka 
Coordinator: Technical Signatory

SABS COMMERCIAL SOC Ltd. Reg. No. 2000/013581/30

Website: www.sabs.co.za E-mail: info@sabs.co.za Call Centre: 086 1277 227



PTS0003

Gauteng Head Office

1 Dr Lategan Road, Groenkloof
Private Bag X191
Pretoria, 0001
Tel +27 (0) 12 428 7911
Fax +27 (0) 12 344 1568

1233